APR - 5 2010

SECTION IV

510(k) SUMMARY OF SAFETY AND EFFECTIVENESS INFORMATION

as required by the Safe Medical Devices Act of 1990 and codified in 21 CFR 807.92 upon which the substantial equivalence is based.

Smith & Nephew FOOTPRINT RCX PK Suture Anchor

Date Prepared: December 18, 2007

A. Submitter's Name:

Smith & Nephew, Inc., Endoscopy Division

150 Minuteman Road

Andover MA, 01810

B. Company Contact

Christina Flores

Regulatory Affairs Specialist II

Phone: (508) 261-3705

FAX: (508) 261-3620

C. Device Name

Trade Name:	FOOTPRINT RCX PK Suture Anchor
Common Name:	Suture Anchor
Classification Name:	Fastener, fixation, non-degradable, soft tissue

D. Predicate Devices

The Smith & Nephew FOOTPRINT RCX PK Anchor is substantially equivalent in Intended Use and Fundamental Scientific Technology to the following legally marketed device in commercial distribution: Smith & Nephew Twinfix PK FP Suture Anchor.

E. Description of Device

The FOOTPRINT RCX PK Suture Anchor is a suture anchor manufactured from PEEK polymer. The screw-in anchor can be loaded with up to four strands of suture and facilitates arthroscopic repair of soft tissue to bone. The design allows

the surgeon to adjust the tension on the tissue intraoperatively and secure the repair.

F. Intended Use

The Smith & Nephew FOOTPRINT RCX PK Suture Anchor is intended for use for the reattachment of soft tissue to bone for the following indications:

Shoulder: Rotator cuff repair, Bankart repair. Slap lesion repair, Biceps tenodesis, Acromioclavicular separation, Deltoid repair, and Capsular shift or Capsulolabral reconstruction.

Foot/Ankle: Lateral stabilization, Medial stabilization, Achilles tendon repair, Hallux valgus reconstruction, Mid-foot reconstruction, Metatarsal ligament repair.

Knee: Medial collateral ligament repair, Lateral collateral ligament repair, Patellar tendon repair, Posterior oblique ligament repair, Illiotibial band tenodesis.

Hand/Wrist: Scapholunate ligament reconstruction, Ulnar collateral ligament reconstruction, Radial collateral ligament reconstruction.

Elbow: Biceps Tendon reattachment, Ulnar or radial collateral ligament reconstruction

G. Comparison of Technological Characteristics

The Smith & Nephew FOOTPRINT RCX PK Suture Anchor is substantially equivalent in materials, fundamental scientific technology, and intended use and is as safe and effective as its currently marketed predicate device, the Smith & Nephew Twinfix PK FP (K073509).

H. Summary Performance Data

The performance testing conducted demonstrates that the insertion and fixation properties of the Smith & Nephew FOOTPRINT RCX PK suture anchor is substantially equivalent to the predicate Smith & Nephew Twinfix PK FP suture anchor, cleared via K073509. The testing also demonstrates the new device does not raise any new issues of safety and efficacy.

DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room –WO66-G609
Silver Spring, MD 20993-0002

Smith & Nephew, Inc.
% Ms. Christina Flores
Regulatory Affairs Specialist II
150 Minuteman Road
Andover, Massachusetts 01810

APR - 5 2010

Re: K093935
Trade/Device Name: Smith & Nephew FOOTPRINT RCX PK Suture Anchor
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth or threaded metallic bone fixation fastener
Regulatory Class: Class II
Product Code: MBI
Dated: March 26, 2010
Received: March 29, 2010

Dear Ms. Flores:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,



Mark N. Melkerson
Director Division of Surgical, Orthopedic
and Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K093935

Device Name: Smith & Nephew FOOTPRINT RCX PK Suture Anchor

Indications For Use:

The Smith & Nephew FOOTPRINT RCX PK Suture Anchor is intended for use for the reattachment of soft tissue to bone for the following indications:

Shoulder: Rotator cuff repair, Bankart repair. Slap lesion repair, Biceps tenodesis, Acromio-Clavicular separation, Deltoid repair, and Capsular shift or Capsulolabral reconstruction.

Foot/Ankle: Lateral stabilization, Medial stabilization, Achilles tendon repair, Hallux valgus reconstruction, Mid-foot reconstruction, Metatarsal ligament repair.

Knee: Medial collateral ligament repair, Lateral collateral ligament repair, Patellar tendon repair, Posterior oblique ligament repair, Illiotibial band tenodesis.

Hand/Wrist: Scapholunate ligament reconstruction, Ulnar collateral ligament reconstruction, Radial collateral ligament reconstruction.

Elbow: Biceps Tendon reattachment, Ulnar or radial collateral ligament reconstruction.

Prescription Use ___x____ AND/OR Over-The-Counter Use _____

(Per 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

(Division Sign-Off)
Division of Surgical, Orthopedic,
and Restorative Devices

510(k) Number K093935